

CERTIFICATE OF ANALYSIS

Product: **WorkBeads™ 40 IDA**

Article number: **40 601 001, 40 601 003, 40 601 010, 40 601 050**

Lot number: **151488**

Product release date: **2016 FEB 26**

Expire date: **2021-02**

	Measured values	Specifications
Capacity:	52.9 $\mu\text{mol Cu}^{2+}/\text{mL}$	50 - 70 $\mu\text{mol Cu}^{2+}/\text{mL}$
Average particle size (Dv50):	51.6 μm	32 - 60 μm
Flow properties (linear flow in a 10x300 mm column):	>10 cm/min	≥ 10 cm/min

Issued 2019 OCT 31
by



Lars Haneskog
Director Quality control