

---

## CERTIFICATE OF ANALYSIS

---

**Product:** BabyBio Zn-IDA 1 ml

Article number: 45 655 041, 45 655 042, 45 655 043, 45 655 044

Lot number: 151524

Resin lot number: 151417

Product release date: 2015 DEC 04

Expire date: 2020-12

---

	Measured values	Specifications
Capacity before charging with Zn <sup>2+</sup>	67 µmol Cu <sup>2+</sup> /ml	50 – 70 µmol Cu <sup>2+</sup> /ml
Gap control: (5 ml/min, water, 22 °C)	Passed	No gap
Flow test: (7 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

---

Issued **2020 MAY 07**

by



---

Mats Johnson  
Chief Executive officer