

CERTIFICATE OF ANALYSIS

Product:	BabyBio Cu-IDA 1 ml
Article number:	45 655 021, 45 655 022, 45 655 023, 45 655 024
Lot number:	151526
Resin lot number:	151422
Product release date:	2015 DEC 07
Expire date:	2020-12

	Measured values	Specifications
Capacity before charging with Cu ²⁺	67 µmol Cu ²⁺ /ml	50 – 70 µmol Cu ²⁺ /ml
Gap control: (5 ml/min, water, 22 °C)	Passed	No gap
Flow test: (7 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

Issued **2020 MAY 07**

by



Mats Johnson
Chief Executive officer