

CERTIFICATE OF ANALYSIS

Product: **WorkBeads™ 40 DEAE**

Article number: **40 150 000**

Lot number: **151588**

Manufacturing date: **2016-04-28**

| | Measured values | Specifications |
|--|------------------------|-----------------------|
| Ionic capacity (mmol Cl ⁻ /mL) | 0.14 | 0.11-0.16 |
| Average particle size (d ₅₀) | 46.9 µm | 32-60 µm |
| Flow properties (linear flow in a 10x300 mm column) | ≥10 cm/min | ≥10 cm/min |

Issued 2017 August 30th

by



Allan Simpson
Quality Assurance Manager