

CERTIFICATE OF ANALYSIS

Product: **WorkBeads™ 40 SEC**

Article number: **40 300 000**

Lot number: **151702**

	Measured values	Specifications
Average particle size (d ₅₀)	53.1 µm	32-60 µm
Kd (Ovalbumin)	0.64	0.62-0.75
Flow properties (linear flow in a 10x300 mm column)	≥10 cm/min	≥10 cm/min

Issued 2017 August 3rd

by



Allan Simpson
Quality Assurance Manager