

## CERTIFICATE OF ANALYSIS

---

**Product:** **BabyBio ACT 5 mL**

Article number: **45 400 005, 45 400 006, 45 400 007, 45 400 008**

Lot number: **151736**

Resin lot number: **BD018**

---

	<b>Measured values</b>	<b>Specifications</b>
Conc. of reactive groups:	252 µmol/mL	> 200 µmol/mL
Gap control: (20 ml/min, water, 22 °C)	Passed	No gap
Flow test: (25 mL/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

---

Issued 2018 JUNE 15

by



---

Mats Johnson  
Chief Executive officer