

---

## CERTIFICATE OF ANALYSIS

---

**Product:** **BabyBio Co-IDA 1 ml**

Article number: **45 655 031, 45 655 032, 45 655 033, 45 655 034**

Lot number: **151750**

Resin lot number: **151401**

Product release date: **2016 FEB 29**

Expire date: **2021-02**

---

	<b>Measured values</b>	<b>Specifications</b>
Capacity before charging with Co <sup>2+</sup>	67 µmol Cu <sup>2+</sup> /ml	50 – 70 µmol Cu <sup>2+</sup> /ml
Gap control: (5 ml/min, water, 22 °C)	Passed	No gap
Flow test: (7 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

---

Issued **2020 MAY 07**

by



---

Mats Johnson  
Chief Executive officer