

CERTIFICATE OF ANALYSIS

Product:	BabyBio Co-IDA 5 ml
Article number:	45 655 035, 45 655 036, 45 655 037, 45 655 038
Lot number:	151753
Resin lot number:	151401
Product release date:	2016 FEB 29
Expire date:	2021-02

	Measured values	Specifications
Capacity before charging with Co ²⁺	67 µmol Cu ²⁺ /ml	50 – 70 µmol Cu ²⁺ /ml
Gap control: (20 ml/min, water, 22 °C)	Passed	No gap
Flow test: (25 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

Issued **2020 JUN 09**

by



Helena Pahlm
Quality Assurance Manager