
CERTIFICATE OF ANALYSIS

Product:	BabyBio Cu-IDA 5 ml
Article number:	45 655 025, 45 655 026, 45 655 027, 45 655 028
Lot number:	151787
Resin lot number:	151420
Product release date:	2015 DEC 07
Expire date:	2020-12

	Measured values	Specifications
Capacity before charging with Cu ²⁺	67 µmol Cu ²⁺ /ml	50 – 70 µmol Cu ²⁺ /ml
Gap control: (20 ml/min, water, 22 °C)	Passed	No gap
Flow test: (25 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

Issued **2020 JUN 09**

by



Helena Pahlm
Quality Assurance Manager