
CERTIFICATE OF ANALYSIS

Product:	BabyBio Ni-IDA 5 ml
Article number:	45 655 005, 45 655 006, 45 655 007, 45 655 008
Lot number:	151790
Resin lot number:	151382
Product release date:	2017 OCT 31
Expire date:	2022-10

	Measured values	Specifications
Capacity before charging with Ni ²⁺	58 µmol Cu ²⁺ /ml	50 – 70 µmol Cu ²⁺ /ml
Gap control: (20 ml/min, water, 22 °C)	Passed	No gap
Flow test: (25 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

Issued **2020 JUN 09**

by



Helena Pahlm
Quality Assurance Manager