

---

## CERTIFICATE OF ANALYSIS

---

**Product:** **BabyBio Ni-IDA 1 ml**

Article number: **45 655 001, 45 655 002, 45 655 003, 45 655 004**

Lot number: **151810**

Resin lot number: **151389**

Product release date: **2016 JAN 18**

Expire date: **2021-01**


---

	<b>Measured values</b>	<b>Specifications</b>
Capacity before charging with Ni <sup>2+</sup>	58 µmol Cu <sup>2+</sup> /ml	50 – 70 µmol Cu <sup>2+</sup> /ml
Gap control: (5 ml/min, water, 22 °C)	Passed	No gap
Flow test: (7 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

---

Issued **2020 MAY 07**

by



---

**Mats Johnson**  
Chief Executive officer