

---

## CERTIFICATE OF ANALYSIS

---

**Product:** BabyBio NTA 5 ml

Article number: 45 655 115, 45 655 116, 45 655 117, 45 655 118

Lot number: 151966

Resin lot number: 600064-3014-034

Product release date: 2015 DEC 31

Expire date: 2020-12


---

	Measured values	Specifications
Metal ion capacity (Tested with Cu <sup>2+</sup> )	24 µmol/ml	20 – 30 µmol/ml
Gap control: (20 ml/min, water, 22 °C)	Passed	No gap
Flow test: (25 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

---

Issued **2020 MAY 29**

by



Helena Pahlm  
Quality Assurance Manager