
CERTIFICATE OF ANALYSIS

Product: BabyBio NTA 1 ml

Article number: 45 655 111, 45 655 112, 45 655 113, 45 655 114

Lot number: 152028

Resin lot number: 60064-3014-031

Product release date: 2016 SEP 30

Expire date: 2021-09

	Measured values	Specifications
Colour:	Passed	White to off-white
Metal ion capacity (Tested with Cu ²⁺)	24 µmol/ml	20 – 30 µmol/ml
Gap control: (5 ml/min, water, 22 °C)	Passed	No gap
Flow test: (7 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

Issued **2020 MAY 28**

by



Helena Pahlm
Quality Assurance Manager