
CERTIFICATE OF ANALYSIS

Product: BabyBio Co-NTA 1 ml

Article number: 45 655 131, 45 655 132, 45 655 133, 45 655 134

Lot number: 152062


Resin lot number: 151404

Product release date: 2016 JAN 08

Expire date: 2021-01

	Measured values	Specifications
Colour:	Passed	Dark pink to purple
Metal ion capacity (Tested with Cu ²⁺)	24.0 µmol/ml	20 – 30 µmol/ml
Gap control: (5 ml/min, water, 22 °C)	Passed	No gap
Flow test: (7 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

Issued 2020 MAR 09
by



Helena Pahlm
Quality Assurance Manager