


CERTIFICATE OF ANALYSIS

Product:	BabyBio Cu-NTA 5 ml
Article number:	45 655 125, 45 655 126, 45 655 127, 45 655 128
Lot number:	152154
Resin lot number:	152145
Product release date:	2017 MAR 22
Expire date:	2022-03

	Measured values	Specifications
Colour:	Passed	Turquoise to light blue
Metal ion capacity (Tested with Cu ²⁺)	22.5 µmol/ml	20 – 30 µmol/ml
Gap control: (20 ml/min, water, 22 °C)	Passed	No gap
Flow test: (25 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

Issued **2020 JUN 04**

by



Helena Pahlm
Quality Assurance Manager