

CERTIFICATE OF ANALYSIS

Product: **WorkBeads 40/100 SEC**

Article number: **40 340 000**

Lot number: **152169**

	<u>Measured values</u>	<u>Specifications</u>
Average particle size d_{50} :	49,9 μm	32-60 μm
K_{av} : Ovalbumin	0,56	0.40 - 0.60
Flow properties (linear flow in a 8x300 mm column):	≥ 10 cm/min	≥ 10 cm/min

Issued 2017 Dec 20
by



Helena Pahlm
Quality Assurance Manager