

---

## CERTIFICATE OF ANALYSIS

---

<b>Product:</b>	<b>BabyBio Zn-NTA 5 ml</b>
Article number:	<b>45 655 145, 45 655 146, 45 655 147, 45 655 148</b>
Lot number:	<b>152172</b>
Resin lot number:	<b>152153</b>
Product release date:	<b>2017 MAR 21</b>
Expire date:	<b>2022-03</b>


---

	<b>Measured values</b>	<b>Specifications</b>
Colour:	Passed	White to off-white
Metal ion capacity (Tested with Cu <sup>2+</sup> )	23 µmol/ml	20 – 30 µmol/ml
Gap control: (20 ml/min, water, 22 °C)	Passed	No gap
Flow test: (25 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

---

Issued **2020 JUN 09**

by



Helena Pahlm  
Quality Assurance Manager