

## CERTIFICATE OF ANALYSIS

---

Product: **WorkBeads 40/100 SEC**

Article number: **40 340 000**

Lot number: **152186**

	<u>Measured values</u>	<u>Specifications</u>
Average particle size $d_{50}$ :	52,2 $\mu\text{m}$	32-60 $\mu\text{m}$
$K_{av}$ : Ovalbumin	0,52	0.40 - 0.60
Flow properties (linear flow in a 8x300 mm column):	$\geq 10$ cm/min	$\geq 10$ cm/min

Issued 2017 Dec 20  
by



---

Helena Pahlm  
Quality Assurance Manager