

CERTIFICATE OF ANALYSIS

Product:	BabyBio S 5 ml
Article number:	45 200 105, 45 200 106, 45 200 107, 45 200 108
Lot number:	152212
Resin lot number:	150407-1
Product release date:	2015 APR 14
Expire date:	2020-04

	Measured values	Specifications
Ionic capacity:	200 $\mu\text{mol Na}^+/\text{ml}$	180-250 $\mu\text{mol Na}^+/\text{ml}$
Gap control: (20 ml/min, water, 22 °C)	Passed	No gap
Maximum flow test: (25 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

Issued 2020 JAN 14

by



Lars Haneskog
Director Quality Control