

CERTIFICATE OF ANALYSIS

Product:	BabyBio Co-NTA 5 mL
Storage solution:	20% ethanol
Article number:	45 655 135, 45 655 136, 45 655 137, 45 655 138
Lot number:	152239
Resin lot number:	152196
Manufacturing date:	2017 JUN 13
Product release date:	2017 JUN 29
Expire date:	2022-05

	Measured values	Specifications
Colour:	Passed	Dark pink to purple
Metal ion capacity (Tested with Cu ²⁺)	22,5 µmol/mL	20 – 30 µmol/mL
Gap control: (20 mL/min, water, 22 °C)	Passed	No gap
Flow test: (25 mL/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

This product is manufactured in Uppsala, Sweden. Country of origin is Sweden.

Issued **2021 JUN 30**

by



Susann Sandström
Quality Assurance Manager