

## CERTIFICATE OF ANALYSIS

---

Product: **WorkBeads™ 40/10 000 SEC**

Article number: **40 350 001, 40 350 003, 40 350 010, 40 350 050,  
40 350 203, 40 350 210, 40 350 250, 40 350 260**

Lot number: **152249**

Product release date: **2017 JUN 16**

Expire date: **2022-06**

---

	Measured values	Specifications
Average particle size (Dv50):	47 µm	32 - 60 µm
Kd (Ovalbumin):	0.75	0.65 - 0.85
Flow properties (linear flow in a 10x300 mm column):	>10 cm/min	≥10 cm/min

---

Issued **2020 MAR 03**

by



Helena Pahlm  
Quality Assurance Manager