
CERTIFICATE OF ANALYSIS

Product: **OptioBio™ 40S 10x100**

Article number: **55 420 011**

Column ID: **152259-02**

Resin lot number: **BD009**

Product release date: **2016 APR 18**

Expire date: **2021-04**

	Measured values	Specifications
Efficiency	14271 N/m	> 7400 N/m
Asymmetry	1.17	0.75 - 1.50
Column backpressure at max recommended flow (450 cm/h, 6 ml/min, water)	0.13 MPa	Not specified
Bed height	102 mm	100 - 102 mm

Issued **2020 MAY 11**

by



Mats Johnson
Chief Executive officer