

CERTIFICATE OF ANALYSIS

Product: **WorkBeads™ 40/1000 SEC**

Article number: **40 300 001, 40 300 003, 40 300 010, 40 300 050**

Lot number: **152364**

Manufacturing date: **2017 SEPTEMBER 14**

Expire Date: **2022-09**

	Measured values	Specifications
Average particle size (Dv50)	49.7 µm	32 - 60 µm
Kd (Ovalbumin)	0.63	0.62 - 0.75
Flow properties (linear flow in a 10x300 mm column)	>10 cm/min	≥ 10 cm/min

Issued 2020 July 14
by



Helena Pahlm
Sr Director Global Quality