

CERTIFICATE OF ANALYSIS

Product:	BabyBio S 5 ml
Article number:	45 200 105, 45 200 106, 45 200 107, 45 200 108
Lot number:	152414
Resin lot number:	152258
Product release date:	2017 JUL 06
Expire date:	2022-07

	Measured values	Specifications
Ionic capacity:	213 $\mu\text{mol Na}^+/\text{ml}$	180-250 $\mu\text{mol Na}^+/\text{ml}$
Gap control: (20 ml/min, water, 22 °C)	Passed	No gap
Maximum flow test: (25 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

Issued 2020 MAR 05

by



Mats Johnson
Chief Executive officer