
CERTIFICATE OF ANALYSIS

| | |
|-----------------------|---|
| Product: | BabyBio Ni-NTA 5 ml |
| Article number: | 45 655 105, 45 655 106, 45 655 107, 45 655 108, 45 655 109 |
| Lot number: | 152435 |
| Resin lot number: | 600064-3020-025 |
| Product release date: | 2017 JUL 30 |
| Expire date: | 2022-07 |

| | Measured values | Specifications |
|---|------------------------|-----------------------|
| Metal ion capacity (Tested with Cu ²⁺) | 23 µmol/ml | 20 – 30 µmol/ml |
| Gap control: (20 ml/min, water, 22 °C) | Passed | No gap |
| Flow test: (25 ml/min, water, 22 °C) | Passed | Stable pressure |
| Visual control: | Passed | No visible defects |

Issued **2020 JUN 04**

by



Helena Pahlm
Quality Assurance Manager