
CERTIFICATE OF ANALYSIS

Product: BabyBio IDA His-tag Screening kit 5 ml

Article number: 45 700 002

Lot number: 152455

Resin lot numbers:

BabyBio Ni-IDA 5ml: 151790

BabyBio Cu-IDA 5ml: 151787

BabyBio Co-IDA 5ml: 151753

BabyBio Zn-IDA 5ml: 151525

Product release date: 2015 JUL 07

Expire date: 2020-07

	Measured values	Specifications
Metal ion capacity (Tested with Cu ²⁺)		50 – 70 µmol/ml
BabyBio Ni-IDA 5 ml	58 µmol/ml	
BabyBio Cu-IDA 5 ml	67 µmol/ml	
BabyBio Co-IDA 5 ml	67 µmol/ml	
BabyBio Zn-IDA 5 ml	67 µmol/ml	
Gap control: (20 ml/min, water, 22 °C)	Passed	No gap
Flow test: (25 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

Issued **2020 MAY 28**

by



Helena Pahlm
Quality Assurance Manager