
CERTIFICATE OF ANALYSIS

Product: BabyBio Ni-NTA 1 ml

Article number: 45 655 101, 45 655 102, 45 655 103, 45 655 104,
45 655 110

Lot number: 152466

Resin lot number: 600064-3020-025

Product release date: 2017 JUL 30

Expire date: 2022-07

	Measured values	Specifications
Colour:	Passed	Turquoise to light blue
Metal ion capacity (Tested with Cu ²⁺)	23 µmol/ml	20 – 30 µmol/ml
Gap control: (5 ml/min, water, 22 °C)	Passed	No gap
Flow test: (7 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

Issued **2020 MAY 07**

by



Mats Johnson
Chief Executive officer