
CERTIFICATE OF ANALYSIS

Product: BabyBio DEAE 5 ml

Article number: 45 150 105, 45 150 106, 45 150 107, 45 150 108

Lot number: 152487

Resin lot number: 151588


Product release date: 2016 JUN 20

Expire date: 2021-06

	Measured values	Specifications
Ionic capacity:	140 $\mu\text{mol Cl}^-/\text{ml}$	110-160 $\mu\text{mol Cl}^-/\text{ml}$
Gap control: (20 ml/min, water, 22 °C)	Passed	No gap
Maximum flow test: (25 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

Issued 2020 FEB 17

by



Lars Haneskog
Director Quality Control