

---

## CERTIFICATE OF ANALYSIS

---

<b>Product:</b>	<b>BabyBio TREN 5 ml</b>
Article number:	<b>45 655 215, 45 655 216, 45 655 217, 45 655 218</b>
Lot number:	<b>152489</b>
Resin lot number:	<b>151845</b>
Product release date:	<b>2016 SEP 13</b>
Expire date:	<b>2021-09</b>

---

	<b>Measured values</b>	<b>Specifications</b>
Metal ion capacity (Tested with Cu <sup>2+</sup> )	59 µmol Cu <sup>2+</sup> /ml	50 – 60 µmol Cu <sup>2+</sup> /ml
Gap control: (20 ml/min, water, 22 °C)	Passed	No gap
Flow test: (25 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

---

Issued **2020 MAY 28**

by



Helena Pahlm  
Quality Assurance Manager