
CERTIFICATE OF ANALYSIS

Product: **BabyBio Q 5 ml**

Article number: **45 100 105, 45 100 106, 45 100 107, 45 100 108**

Lot number: **152490**

Resin lot number: **151946**

Product release date: **2016 NOV 25**

Expire date: **2021-11**

	Measured values	Specifications
Ionic capacity:	180 $\mu\text{mol Cl}^-/\text{ml}$	180-250 $\mu\text{mol Cl}^-/\text{ml}$
Gap control: (20 ml/min, water, 22 °C)	Passed	No gap
Maximum flow test: (25 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

Issued 2020 MAR 05

by



Mats Johnson
Chief Executive officer