
CERTIFICATE OF ANALYSIS

Product: BabyBio Q 1 ml
Article number: 45 100 101, 45 100 102, 45 100 103, 45 100 104
Lot number: 152500
Resin lot number: 151946
Product release date: 2016 AUG 25
Expire date: 2021-08

	Measured values	Specifications
Ionic capacity	180 µmol Cl ⁻ /ml	180-250 µmol Cl ⁻ /ml
Gap control: (5 mL/min, water, 22 °C)	Passed	No gap
Maximum flow test: (7 mL/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

Issued **2020 JUN 09**

by



Helena Pahlm
Quality Assurance Manager