

---

## CERTIFICATE OF ANALYSIS

---

**Product:** BabyBio Co-NTA 5 mL

Article number: 45 655 135, 45 655 136, 45 655 137, 45 655 138

Lot number: 152503

Resin lot number: 152272

Product release date: 2017 AUG 10

Expire date: 2022-08

---

	Measured values	Specifications
Colour:	Passed	Dark pink to purple
Metal ion capacity (Tested with Cu <sup>2+</sup> )	23 µmol/mL	20 – 30 µmol/mL
Gap control: (20 mL/min, water, 22 °C)	Passed	No gap
Flow test: (25 mL/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

---

Issued **2020 JUN 09**

by



Helena Pahlm  
Quality Assurance Manager