
CERTIFICATE OF ANALYSIS

Product:	BabyBio DEAE 1 ml
Article number:	45 150 101, 45 150 102, 45 150 103, 45 150 104
Lot number:	152531
Resin lot number:	151588
Product release date:	2016 JUN 20
Expire date:	2021-06

	Measured values	Specifications
Ionic capacity:	140 $\mu\text{mol Cl}^-/\text{ml}$	110-160 $\mu\text{mol Cl}^-/\text{ml}$
Gap control: (5 ml/min, water, 22 °C)	Passed	No gap
Maximum flow test: (7 mL/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

Issued **2020 MAY 28**

by



Helena Pahlm
Quality Assurance Manager