

---

## CERTIFICATE OF ANALYSIS

---

<b>Product:</b>	<b>BabyBio Ni-NTA 5 ml</b>
Article number:	<b>45 655 105, 45 655 106, 45 655 107, 45 655 108, 45 655 109</b>
Lot number:	<b>152545</b>
Resin lot number:	<b>600064-3020-034</b>
Product release date:	<b>2017 DEC 12</b>
Expire date:	<b>2022-12</b>


---

	<b>Measured values</b>	<b>Specifications</b>
Metal ion capacity (Tested with Cu <sup>2+</sup> )	24.0 µmol/ml	20 – 30 µmol/ml
Gap control: (20 ml/min, water, 22 °C)	Passed	No gap
Flow test: (25 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

---

Issued 2020 MAR 09

by



---

Helena Pahlm  
Quality Assurance Manager