
CERTIFICATE OF ANALYSIS

Product: **WorkBeads™ 100S**

Article number: **10 200 001, 10 200 002, 10 200 005, 10 200 010,
10 200 050, 10 200 060**

Lot number: **152566**

Product release date: **2019 MAY 09**

Expire date: **2024-05**

	Measured values	Specifications
Ion exchange capacity:	210 µmol Na ⁺ /ml	180 - 250 µmol Na ⁺ /ml
Average particle size (Dv50):	100 µm	90 -110 µm
Flow acceptance (linear flow in a 25x200 mm column):	1050 cm/h	≥1000 cm/h

Issued 2019 JUNE 27

by



Lars Haneskog
Director Quality control