

---

## CERTIFICATE OF ANALYSIS

---

**Product:** BabyBio S 1 ml

Article number: 45 200 101, 45 200 102, 45 200 103, 45 200 104

Lot number: 152651

Resin lot number: 152258

Product release date: 2017 JUL 06

Expire date: 2022-07

---

	Measured values	Specifications
Ionic capacity	213 $\mu\text{mol Na}^+/\text{ml}$	180-250 $\mu\text{mol Na}^+/\text{ml}$
Gap control: (5 ml/min, water, 22 °C)	Passed	No gap
Maximum flow test: (7 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

---

Issued **2020 MAY 28**

by



Helena Pahlm  
Quality Assurance Manager