
CERTIFICATE OF ANALYSIS

Product:	BabyBio TREN 5 ml
Article number:	45 655 215, 45 655 216, 45 655 217, 45 655 218
Lot number:	152653
Resin lot number:	151845
Product release date:	2016 SEP 13
Expire date:	2021-09

	Measured values	Specifications
Metal ion capacity (Tested with Cu ²⁺)	59 µmol Cu ²⁺ /ml	50 – 60 µmol Cu ²⁺ /ml
Gap control: (20 ml/min, water, 22 °C)	Passed	No gap
Flow test: (25 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

Issued 2020 MAR 05

by



Mats Johnson
Chief Executive officer