

## CERTIFICATE OF ANALYSIS

Product: **OptioBio™ 40Q 10x100**

Article number: **55 410 011**

Column ID: **152670-03**

Resin lot number: **151946**

	Measured values	Specifications
Efficiency (Test method W077)	9272 N/m	> 6350 N/m
Asymmetry (Test method W077)	1.06	0.75 - 1.50
Column backpressure at max recommended flow (Test method W077; 450 cm/h, 6 ml/min, water)	0.12 MPa	Not specified
Bed height	101 mm	100 - 102 mm

Issued 2018 MAY 25  
by



\_\_\_\_\_  
Helena Pahlm  
Sr Director Global Quality