

## CERTIFICATE OF ANALYSIS

Product: **OptioBio™ 40Q 10x100**

Article number: **55 410 011**

Column ID: **152670-06**

Resin lot number: **151946**


Product release date: **2016 OCT 12**

Expire date: **2021-10**

	Measured values	Specifications
Efficiency	9422 N/m	> 6350 N/m
Asymmetry	1.07	0.75 - 1.50
Column backpressure at max recommended flow (450 cm/h, 6 ml/min, water)	0.09 MPa	Not specified
Bed height	102 mm	100 - 102 mm

Issued 2019 MAY 21

by



Lars Haneskog  
Director of Quality Control