

CERTIFICATE OF ANALYSIS


Product: **WorkBeads™ 100S**

Article number: **10 200 061**

Resin lot number: **152693**

	Measured values	Specifications
Ionic capacity	190	180-250 $\mu\text{mol H}^+$ /mL resin
Flow acceptance (in 10 x 300 mm column)	>10	≥ 10 cm/min
Average particle size, $dV_{(50)}$	119	75-250 μm

Issued 2018 March 26th
by



Helena Pahlm
Quality Assurance Manager

CERTIFICATE OF ANALYSIS

This page only for template!

Prepared/ revised by	Signature	Position	Date
	Lars Haneskog	Dir. of R&D	2017-12-12
Reviewed by	Signature	Position	Date
	Kajsa Eriksson	Res. Eng. R&D	2017-12-12
Approved by	Signature	Position	Date
	Helena Pahlm	Dir. of Quality assurance	2017-12-12

Revisions

Version	Changes	Date
1.0	First version	2017-12-12