
CERTIFICATE OF ANALYSIS

Product: BabyBio NTA 5 ml

Article number: 45 655 115, 45 655 116, 45 655 117, 45 655 118

Lot number: 152727


Resin lot number: 600064-3014-031

Product release date: 2016 SEP 30

Expire date: 2021-09

	Measured values	Specifications
Metal ion capacity (Tested with Cu ²⁺)	24.0 µmol/ml	20 – 30 µmol/ml
Gap control: (20 ml/min, water, 22 °C)	Passed	No gap
Flow test: (25 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

Issued 2020 MAR 09
by



Helena Pahlm
Quality Assurance Manager