
CERTIFICATE OF ANALYSIS

Product: **WorkBeads™ 40 Q**

Article number: **40 100 001, 40 100 002, 40 100 010, 40 100 050,
40 100 060, 40 100 210, 40 100 250**

Lot number: **152792**

Manufacturing date: **2018 MAY 17**

	Measured values	Specifications
Ionic capacity	196 µmol Cl ⁻ /mL	180-250 µmol Cl ⁻ /mL
Average particle size (Dv50)	58.5 µm	32-60 µm
Flow properties (linear flow in a 10x300 mm column)	>20 cm/min	≥10 cm/min

Issued 2018 JULY 30

by



Mats Johnson
Chief Executive officer