
CERTIFICATE OF ANALYSIS

Product: BabyBio IDA His-tag Screening kit 1 ml

Article number: 45 700 001

Lot number: 152807

Resin lot numbers:

BabyBio Ni-IDA 1ml: 151810
BabyBio Cu-IDA 1ml: 151526
BabyBio Co-IDA 1ml: 151750
BabyBio Zn-IDA 1ml: 151524

Product release date: 2015 DEC 07
Expire date: 2020-12

	Measured values	Specifications
Metal ion capacity (Tested with Cu ²⁺)		50 – 70 µmol/ml
BabyBio Ni-IDA 1 ml	58 µmol/ml	
BabyBio Cu-IDA 1 ml	67 µmol/ml	
BabyBio Co-IDA 1 ml	67 µmol/ml	
BabyBio Zn-IDA 1 ml	67 µmol/ml	
Gap control: (5 ml/min, water, 22 °C)	Passed	No gap
Flow test: (7 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

Issued **2020 MAY 28**

by



Helena Pahlm
Quality Assurance Manager