
CERTIFICATE OF ANALYSIS

Product: **WorkBeads™ 40 Co-NTA**

Article number: **40 651 401, 40 651 403, 40 651 410, 40 651 411**

Lot number: **152906**

Product release date: **2016 SEP 30**

Expire date: **2021-09**

	Measured values	Specifications
Capacity before charging with Co ²⁺	22.5 µmol Cu ²⁺ /ml	20 - 30 µmol Cu ²⁺ /ml
Average particle size (Dv50):	49.5 µm	32 - 60 µm
Flow acceptance (linear flow in a 10x300 mm column):	>10 cm/min	≥10 cm/min

Issued 2019 JUL 17

by



Lars Haneskog
Director of Quality Control