
CERTIFICATE OF ANALYSIS

Product: **OptioBio™ 40S 10x100**

Article number: **55 420 011**

Column ID: **152908-13**

Resin lot number: **152258**


Product release date: **2017 JUL 06**

Expire date: **2022-07**

	Measured values	Specifications
Efficiency	14499 N/m	> 7400 N/m
Asymmetry	1.10	0.75 - 1.50
Column backpressure at max recommended flow (450 cm/h, 6 ml/min, water)	0.05 MPa	Not specified
Bed height	101 mm	100 - 102 mm

Issued 2020 JAN 17

by



Lars Haneskog
Director of Quality Control