
CERTIFICATE OF ANALYSIS

Product: **OptioBio™ 40S 10x100**

Article number: **55 420 011**

Column ID: **152908-16**

Resin lot number: **152258**


Product release date: **2017 JUL 06**

Expire date: **2022-07**

| | Measured values | Specifications |
|-------------------------------------------------------------------------|-----------------|----------------|
| Efficiency | 13896 N/m | > 7400 N/m |
| Asymmetry | 1.21 | 0.75 - 1.50 |
| Column backpressure at max recommended flow (450 cm/h, 6 ml/min, water) | 0.04 MPa | Not specified |
| Bed height | 101 mm | 100 - 102 mm |

Issued 2020 JAN 17

by



Lars Haneskog
Director of Quality Control