

## CERTIFICATE OF ANALYSIS

---

Product: **WorkBeads™ 40/100 SEC**

Article number: **40 340 001, 40 340 003, 40 340 010, 40 340 050,  
40 340 250, 40 340 260**

Lot number: **152969**

Manufacturing date: **2018 SEPTEMBER 11**

---

	<b>Measured values</b>	<b>Specifications</b>
Average particle size (Dv50):	50.7 µm	32 - 60 µm
Kd (Ovalbumin):	0.55	0.40 - 0.60
Flow properties (linear flow in a 10x300 mm column):	>10 cm/min	≥10 cm/min

---

Issued 2018 SEPTEMBER 12

by



---

**Mats Johnson**  
Chief Executive officer