

CERTIFICATE OF ANALYSIS

Product: **WorkBeads™ 40/100 SEC**

Article number: **40 340 001, 40 340 003, 40 340 010, 40 340 050,
40 340 250, 40 340 260**

Lot number: **152977**

Product release date: **2019 JUL 22**

Expire date: **2024-07**

	Measured values	Specifications
Average particle size (Dv50):	57 µm	32 - 60 µm
Kd (Ovalbumin):	0.45	0.40 - 0.60
Flow properties (linear flow in a 10x300 mm column):	>10 cm/min	≥10 cm/min

Issued 2019 JUL 22

by



Lars Haneskog
Director Quality Control