

---

## CERTIFICATE OF ANALYSIS

---

<b>Product:</b>	<b>BabyBio™ affimAb 1 ml</b>
Article number:	<b>45 800 101, 45 800 102, 45 800 103, 45 800 104</b>
Lot number:	<b>153001</b>
Resin lot number:	<b>152968</b>
Product release date:	<b>2018 SEP 25</b>
Expire date:	<b>2021-09</b>

---

	<b>Measured values</b>	<b>Specifications</b>
Dynamic binding capacity for IgG <sup>1</sup> .	41.0 mg/ml	> 40 mg/ml
Gap control: (4 ml/min, water, 22 °C)	Passed	No gap
Maximum flow test: (7 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

---

1. Test done at 2.5 minutes residence time in 6.6x100 mm column.

Issued 2020 MAR 05

by



---

**Mats Johnson**  
Chief Executive officer