

---

## CERTIFICATE OF ANALYSIS

---

**Product:** BabyBio™ affimAb 5 ml

Article number: 45 800 105, 45 800 106, 45 800 107, 45 800 108

Lot number: 153002

Resin lot number: 152968

Product release date: 2018 OCT 25

Expire date: 2021-10

---

	Measured values	Specifications
Dynamic binding capacity for IgG <sup>1</sup> .	41 mg/ml	> 40 mg/ml
Gap control: (15 ml/min, water, 22 °C)	Passed	No gap
Maximum flow test: (17 ml/min, water, 22 °C)	Passed	Stable pressure
Visual control:	Passed	No visible defects

---

1. Test done at 2.5 minutes residence time in 6.6x100 mm column.

Issued **2020 MAY 07**

by



---

Mats Johnson  
Chief Executive officer